



FACTORY 7/91-99 BERESFORD ROAD
 P O BOX 189 LILYDALE
 VICTORIA AUSTRALIA 3140

www.industrilex.com.au
 Telephone (03) 9735 1673
 Fax (03) 9739 6490

Engineers Work Sheet

Customer

Service Date

Hospital	Nursing Home	Preventative Maintenance	Thermocouple	Incomplete	Completed
Machine Type		Model No.		Serial No.	
Fault/Action			Location		

Tests	Pass	Failed	N/A		Pass	Failed	N/A
Thermocouple	___	___	___	Detergent Dispensing	___	___	___

Residual Detergent test	Pass	Failed	N/A	Ph test	___	Water Hardness	___
	___	___	___	Chloride Residue	___	Chlorine Content	___
				Iron Content	___		

Preventative Maintenance in accordance with Product Manual

Hot water valve	___	Heater	___	Timer	___
Cold water valve	___	Wash Filters	___	Temp. Indicator	___
Wash pump	___	Electrical Comp	___	Acid rinse	___
Drain pump	___	Basket wheels	___	Control Panel	___
Tempreture Control	___	Chamber condition	___	Inlet filters	___
Printer/Chart Recorder	___	Upper venture	___	Printer Paper	___
Water hoses	___	Printer ribbon	___	Door seal	___
Indicators/Gauges	___	Door locking	___		

Parts used

Comments/Recommendations

Service Label _____